



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/606,304	06/25/2003	David Wayne Daniels	DEP5083	8039

27777 7590 12/02/2010  
PHILIP S. JOHNSON  
JOHNSON & JOHNSON  
ONE JOHNSON & JOHNSON PLAZA  
NEW BRUNSWICK, NJ 08933-7003

EXAMINER
----------

HOFFMAN, MARY C

ART UNIT	PAPER NUMBER
----------	--------------

3733

NOTIFICATION DATE	DELIVERY MODE
-------------------	---------------

12/02/2010

ELECTRONIC

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

jnjuspatent@corus.jnj.com  
lhowd@its.jnj.com  
gsanche@its.jnj.com

<b>Office Action Summary</b>	<b>Application No.</b> 10/606,304	<b>Applicant(s)</b> DANIELS ET AL.	
	<b>Examiner</b> MARY HOFFMAN	<b>Art Unit</b> 3733	

**-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --**

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 24 September 2010.
- 2a) ☒ This action is **FINAL**.                      2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-3 and 5-29 is/are pending in the application.
- 4a) Of the above claim(s) 9-29 is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-3, 5-8 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 05 July 2008 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \*    c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |   |   |
|---|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)         | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____                                      |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)         | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____   | 6) <input type="checkbox"/> Other: _____                          |

## DETAILED ACTION

### ***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-3 and 5-8 are rejected under 35 U.S.C. 103(a) as being unpatentable over Shaolian et al. (U.S. Patent No. 7,641,658) in view of Middleton et al. (U.S. Patent Application Publication 2002/0183758).

Shaolian et al. disclose a reamer (see FIGS. 9 and 10) for preparing a cavity in the intramedullary canal a long bone, the reamer comprising a first component (ref. #238) for preparation of the cavity in the canal, the first component including a portion thereof for placement at least partially in the cavity of the long bone, the first component defining a rotational centerline thereof; and a rigid second component (ref. #236) operably connected to the first component, the second component defining a rotational centerline thereof, the rotational centerline of the first component and the rotational centerline of the second component having a first relationship in which the centerlines are coincident and a second relationship in which the centerlines are skewed with respect to each other (see FIG. 10). The reamer includes a rotatable locking arrangement (ref. #246) by which the first and second components can be locked against relative movement from the second to the first relationship. The reamer further

Art Unit: 3733

comprises a joint (ref. #240) operably connected to the first component and to the second component, the joint adapted to provide the first relationship in which the centerlines are coincident and the second relationship in which the centerlines are skewed with respect to each other. The locking arrangement includes a pin (ref. #240). The first component includes a portion thereof having a tapered external periphery and wherein the second component includes a portion thereof having a drive connection. The reamer can be used in a femur. The first component and the second component are hinged to each other (at ref. #240). The first component and the second component define openings therein for receiving a pin (at ref. #240).

Shaolian et al. discloses the claimed invention except for the first component including a plurality of flutes (rather, Shaolian et al. disclose a blade-like first component).

Middleton et al. discloses a blade-like first component (see ref. #320, FIG. 5A) that comprises a cutting tip (ref. #330) including cutting flutes (ref. #360) to aid in cutting the tissue.

It would have been obvious to one of ordinary skill in the art at the time the invention was made to modify the blade-like first component of Shaolian et al. with a cutting tip including cutting flutes in view of Middleton et al. in order to aid in cutting the tissue.

***Response to Arguments***

Applicant's arguments with respect to the claims have been considered but are moot in view of the new ground(s) of rejection.

***Conclusion***

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to MARY HOFFMAN whose telephone number is (571)272-5566. The examiner can normally be reached on Monday-Thursday 10:00-5:00pm.

Art Unit: 3733

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo C. Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Mary C. Hoffman/  
Examiner, Art Unit 3733

/EDUARDO C. ROBERT/

Supervisory Patent Examiner, Art Unit 3733